



## Ohio Administrative Code

Rule 3701:1-58-23 Possession, use, and calibration of instruments used to measure the activity of unsealed radioactive material.

Effective: August 15, 2005

---

(A) For direct measurements performed in accordance with rule 3701:1-58-25 of the Administrative Code, a licensee shall possess and use instrumentation to measure the activity of unsealed radioactive material before it is administered to each patient or human research subject.

(B) A licensee shall calibrate the instrumentation required in paragraph (A) of this rule in accordance with nationally recognized standards or the manufacturer's instructions.

(C) A licensee shall retain a record of each instrument calibration required by this rule in accordance with rule 3701:1-58-77 of the Administrative Code.

---